



RECEIVED

SEP 25 2003

TECHNOLOGY CENTER R3700

EXHIBIT A

Copy of Spec.

-1-

METHOD FOR ACQUIRING MRI DATA FROM A LARGE FIELD
OF VIEW USING CONTINUOUS TABLE MOTION

STATEMENT REGARDING FEDERALLY SPONSORED RESEARCH

[0001] This invention was made with government support under Grant Nos. CA37993 and HL37310 awarded by the National Institute of Health. The United States Government has certain rights in this invention.

CROSS-REFERENCE TO RELATED APPLICATIONS

[0002] This application is based on U.S. Provisional Patent Application Serial No. 60/282,555 filed on April 9, 2001 and entitled "Method For Acquiring MRI Data From A Large Field Of View Using Continuous Table Motion."

BACKGROUND OF THE INVENTION

[0003][0002]—The field of the invention is magnetic resonance angiography ("MRA"), and particularly, studies of the human vasculature using contrast agents which enhance the NMR signals.

[0004][0003]—Magnetic resonance angiography (MRA) uses the nuclear magnetic resonance (NMR) phenomenon to produce images of the human vasculature. When a substance such as human tissue is subjected to a uniform magnetic field (polarizing field B_0), the individual magnetic moments of the spins in the tissue attempt to align with this polarizing field, but precess about it in random order at their characteristic Larmor frequency. If the substance, or tissue, is subjected to a magnetic field (excitation field B_1) which is in the x-y plane and which is near the Larmor frequency, the net aligned moment, M_z , may be rotated, or "tipped", into the x-y plane to produce a net transverse magnetic moment M_t . A signal is emitted by the excited spins, and after the excitation signal B_1 is terminated, this signal may be received and processed to form an image.

[0005][0004]—When utilizing these signals to produce images, magnetic field gradients (G_x , G_y and G_z) are employed. Typically, the region to be imaged is scanned by a sequence of measurement cycles in which these gradients vary according to the particular localization method being used. The resulting set of

received NMR signals, or "views" are digitized and processed to reconstruct the image using one of many well known reconstruction techniques.

[0006]~~[0005]~~—MR angiography (MRA) is the application of magnetic resonance imaging methods to the depiction of the human vasculature. To enhance the diagnostic capability of MRA a contrast agent such as gadolinium can be injected into the patient prior to the MRA scan. Excellent diagnostic images may be acquired using contrast-enhanced MRA if the data acquisition is properly timed with the bolus passage.

[0007]~~[0006]~~—The non-invasiveness of MRA makes it a valuable screening tool for cardiovascular diseases. Screening typically requires imaging vessels in a large volume. This is particularly true for diseases in the runoff vessels of the lower extremity. The field of view (FOV) in MR imaging is limited by the volume of the B_0 field homogeneity and the receiver coil size (typically, the $FOV < 48$ cm on current commercial MR scanners). The anatomic region of interest in the lower extremity, for example, is about 100 cm and this requires several FOVs, or stations, for a complete study. This requires that the patient be repositioned inside the bore of the magnet, the patient be re-landmarked, scout images be acquired and a preparation scan be performed for each FOV. All of these additional steps take time and, therefore, are expensive. When contrast enhanced MRA is performed, the repositioning also necessitates additional contrast injections.

[0008]~~[0007]~~—Recently gadolinium-enhanced bolus chase techniques have been reported which overcome this difficulty, K.Y. Ho, T. Leiner, M.H. de Hann, J.M.A. van Engleshoven, "Gadolinium optimized tracking technique: a new MRA technique for imaging the peripheral vascular tree from aorta to the foot using one bolus of gadolinium (abs)." Proc. 5th Meeting of ISMRM, p203, 1997. As described in U.S. Pat. Nos. 5,924,987 and 5,928,148, MRA data is acquired from a large field of view by automatically moving the patient table to a plurality of different locations during the scan and acquiring an image at each station. The movement of the table may be timed to follow the contrast bolus through the vasculature so that peak contrast is achieved at each station.

[0009]~~[0008]~~—The result of prior bolus chase MRA methods is that one ends up with a plurality of images. These are manually or automatically registered

with each other to provide a single image that covers the entire extended field of view. One difficulty with this approach, however, is that the separate images have different brightnesses and/or contrasts. As a result, there are discontinuities at the boundaries of images where they have been patched together. Another difficulty with the multi-station method is that valuable time is lost when the table is moved from one station to the next. During that time no image data is being acquired and further time is lost in bringing the spin magnetization into dynamic equilibrium before image data is acquired. In a three-dimensional scan this lost data acquisition time can mean that the scanning process does not keep pace with the moving contrast bolus and some image contrast is lost in later images.

SUMMARY OF THE INVENTION

[0010]~~[0009]~~—The present invention is a method for acquiring an MR image from an extended field of view in which the patient table is continuously moved during the scan. The acquired MRI data is motion corrected to a common table reference position and a single, seamless MR image is reconstructed. MRI data acquired during the scan may be used to reconstruct images in real-time that enable the operator to monitor the progress of the contrast bolus and to adjust the speed of the table motion accordingly to insure peak contrast throughout the scan.

[0011]~~[0010]~~—A general object of the invention is to produce a single MR image over a field of view which exceeds the normal field of view of the MRI system. The subject is moved continuously through the MRI system and views are acquired along with table position data. Each view is position corrected using the associated table position data to produce a single array of MRI data which is used to reconstruct an image.

[0012]~~[0011]~~—Another object of the invention is to better time the acquisition of MRI data with the movement of the contrast bolus through the subject. Once the scan is begun no time is wasted. For a given image resolution and pulse sequence, the table is moved continuously at the maximum possible velocity. Table velocity can be slowed as needed to remain properly aligned with peak contrast as it traverses through the subject.

[0013][0012]—The foregoing and other objects and advantages of the invention will appear from the following description. In the description, reference is made to the accompanying drawings which form a part hereof, and in which there is shown by way of illustration a preferred embodiment of the invention. Such embodiment does not necessarily represent the full scope of the invention, however, and reference is made therefore to the claims herein for interpreting the scope of the invention.

BRIEF DESCRIPTION OF THE DRAWINGS

[0014][0013]—Fig. 1 is a block diagram of an MRI system which employs the present invention;

[0015][0014]—Fig. 2 is a graphic representation of a pulse sequence performed by the MRI system of Fig. 1 to practice a preferred embodiment of the invention;

[0016][0015]—Fig. 3 is a schematic representation of an exemplary imaging volume within the bore of the MRI system of Fig. 1;

[0017][0016]—Fig. 4 is a schematic representation of a data matrix into which data acquired with the pulse sequence of Fig. 2 is stored; and

[0018][0017]—Fig. 5 is a flow chart illustrating the steps employed in practicing the preferred embodiment of the invention.

GENERAL DESCRIPTION OF THE INVENTION

[0019][0018]—The present invention allows the subject to be imaged over a field of view (FOV) which can be many times larger than the static FOV allowed by a typical MRI scanner. The invention allows the MRI scanner to run uninterrupted while the patient table is continuously moved through the portion of the scanner volume which presents the best combination of homogeneous static magnetic field and linear magnetic field gradients. This region is typically known as the scanner's "sweet spot." The invention allows the acquisition of arbitrarily large FOVs along the table motion direction with one uninterrupted data set. Two and three-dimensional imaging are possible. The method of imaging during continuous table motion can be coupled with other techniques, including real-time MR imaging, to provide real-time control of the table motion and scan parameters. The invention is

primarily directed toward MR angiography, but is also a general MRI imaging method.

[0020]~~[0019]~~—The invention can be described by referring to Fig. 3 which schematically shows the bore 10 of an MRI scanner. For this discussion a slab 12 with coronal orientation is being imaged. The readout direction (X axis) is along the direction of motion of the table and the slab select direction is along the Z axis. For 2DFT imaging, phase encoding is performed along the Y axis and for 3DFT imaging, phase encoding is performed along both the Y and Z axes.

[0021]~~[0020]~~—We define FOV_{tot} as the full desired field of view along the X direction and FOV_x as the field of view along X for the readout of any single NMR signal. We assume that $FOV_{tot} > FOV_x$ thus necessitating the invention described here to provide an image for the full FOV if continuous table motion is desired.

[0022]~~[0021]~~—As the subject is moved through the scanner along the x-axis, MRI data is acquired from the slab 12 by a series of imaging pulse sequences. During each pulse sequence the readout of the encoded k-space data is performed in the X direction. The raw k-space data is typically anti-alias filtered and subsequently sampled at a bandwidth assuming some FOV_x in the readout (X) direction.

[0023]~~[0022]~~—Referring to Fig. 4, the raw k-space data are acquired in a series of imaging pulse sequences, a Fourier transformation (FT) is performed on each NMR signal in the readout direction and the result is placed in a matrix of memory locations 16 which represents the desired resolution in the X, Y and Z directions. This once-transformed view is a line of data placed in the matrix 16 and shifted in the X direction to a position that represents the patient position at the time the NMR echo signal was acquired. The position of the patient table relative to a reference location is known for every echo readout. The data acquisition is performed until all of the data matrix 16 is filled and the table has reached the final position. The final image is formed by Fourier transforming along the remaining Y and if 3D, the Z directions.

[0024] The following is a discussion of the theory of signal alteration due to arbitrary positioning of the patient and the adjustment of the acquired NMR echo

signal. The MR signal measured during a standard image acquisition can be described by:

[0023]— ~~The following is a discussion of the theory of signal alteration due to arbitrary positioning of the patient and the adjustment of the acquired NMR echo-signal. The MR signal measured during a standard image acquisition can be described by:~~

$$S_n(t) = \int \int m(x, y) e^{-i\gamma G_n y t_y} e^{-i\gamma G_x x t} dx dy \quad (1)$$

where x is the frequency-encoding (readout) direction, y is the phase-encoding direction, G_x and G_n are the respective gradients, $m(x, y)$ is the transverse magnetization of the object in spatial coordinates, γ is the gyromagnetic ratio, and t_y is the duration of the y gradient. The index n is the repetition number of the acquisition and runs from 0 to N-1 where N is the total number of phase encodings along the Y axis.

[0024]— ~~An object, shifted from its initial reference position along x some arbitrary distance has position~~ An object, shifted from its initial reference position along x some arbitrary distance has position Δ . ~~The signal from this shifted object is then given by:~~

(2)

~~Now by substitution of variables~~

The signal from this shifted object is then given by:

we have

y

$$S_n(t) = \iint m(x - \Delta, y) e^{-i\gamma G_n y t_y} e^{-i\gamma G_x x t} dx dy \quad (2)$$

Now by substitution of variables

$$x' = x - \Delta \quad \text{and} \quad dx' = dx$$

we have

$$S_n(t) = \iint m(x', y) e^{-i\gamma G_n y t_y} e^{-i\gamma G_x (x' + \Delta) t} dx' dy \quad (3)$$

and

and

$$S_n(t) = e^{-i\gamma G_x \Delta t} \iint m(x', y) e^{-i\gamma G_n y t_y} e^{-i\gamma G_x x' t} dx' dy \quad (4)$$

(4)

where $e^{-i\gamma G_x \Delta t}$ is a phase term representing the positional shift of the object in k-space. It is assumed that γ, γ_n , G_n and Δ are known for any specific phase encoding.

The phase factor required to recover or unshift the signal $S_n(t)$ is then $e^{+i\gamma G_x \Delta t}$.

[0025] [0025]—There are two possible methods to position each echo using the previous theory. The direct method in which the shifted data is repositioned by applying the phase factor to the continuous presampled, raw k-space data, and the time and memory efficient method in which a combination of phase shifting and

positional shifting is applied to the sampled data. These are described in more detail in the description of the preferred embodiment of the invention.

DESCRIPTION OF THE PREFERRED EMBODIMENT

[0026]—Referring first to Fig. 1, there is shown the major components of a preferred MRI system which incorporates the present invention. The operation of the system is controlled from an operator console 100 which includes a keyboard and control panel 102 and a display 104. The console 100 communicates through a link 116 with a separate computer system 107 that enables an operator to control the production and display of images on the screen 104. The computer system 107 includes a number of modules which communicate with each other through a backplane. These include an image processor module 106, a CPU module 108 and a memory module 113, known in the art as a frame buffer for storing image data arrays. The computer system 107 is linked to a disk storage 111 and a tape drive 112 for storage of image data and programs, and it communicates with a separate system control 122 through a high speed serial link 115.

[0027]—The system control 122 includes a set of modules connected together by a backplane. These include a CPU module 119 and a pulse generator module 121 which connects to the operator console 100 through a serial link 125. It is through this link 125 that the system control 122 receives commands from the operator which indicate the scan sequence that is to be performed. The pulse generator module 121 operates the system components to carry out the desired scan sequence. It produces data which indicates the timing, strength and shape of the RF pulses which are to be produced, and the timing of and length of the data acquisition window. The pulse generator module 121 connects to a set of gradient amplifiers 127, to indicate the timing and shape of the gradient pulses to be produced during the scan. The pulse generator module 121 also receives patient data from a physiological acquisition controller 129 that receives signals from a number of different sensors connected to the patient, such as ECG signals from electrodes or respiratory signals from a bellows. And finally, the pulse generator module 121 connects to a scan room interface circuit 133 which receives signals from various sensors associated with the condition of the patient and the magnet

system. It is also through the scan room interface circuit 133 that a patient positioning system 134 receives commands from the pulse generator module 121 to move the patient through the scanner to perform the scan in accordance with the present invention. The current position of the table at any time during the scan is read into the system control 122 and is used to adjust the acquired NMR data according to the present invention as will be described in more detail below. The operator can control the operation of the patient positioning system 134 through the keyboard and control panel 102. This includes controlling the velocity of table motion during the scan.

[0028]—The gradient waveforms produced by the pulse generator module 121 are applied to a gradient amplifier system 127 comprised of G_x , G_y and G_z amplifiers. Each gradient amplifier excites a corresponding gradient coil in an assembly generally designated 139 to produce the magnetic field gradients used for position encoding acquired signals. The gradient coil assembly 139 forms part of a magnet assembly 141 which includes a polarizing magnet 140 and a whole-body RF coil 152. A transceiver module 150 in the system control 122 produces pulses which are amplified by an RF amplifier 151 and coupled to the RF coil 152 by a transmit/receive switch 154. The resulting signals radiated by the excited nuclei in the patient may be sensed by the same RF coil 152 and coupled through the transmit/receive switch 154 to a preamplifier 153. The amplified NMR signals are demodulated, filtered, and digitized in the receiver section of the transceiver 150.

~~**[0029]**— The transmit/receive switch 154 is controlled by a signal from the pulse generator module 121 to electrically connect the RF amplifier 151 to the coil 152 during the transmit mode and to connect the preamplifier 153 during the receive mode. The transmit/receive switch 154 also enables a separate RF local coil to be used during the receive mode. In the preferred embodiment of the invention a single, stationary local coil described below and shown in Fig. 7 is switched into operation. In the alternative, a multi-segment local coil described below and shown in Fig. 8 may also be used. In this case, the separate segments of the local coil are switched into operation by the transmit/receive switch 154 as the patient is translated to successive stations during the scan.~~

[0029] The NMR signals picked up by the RF local coil are digitized by the transceiver module 150 and transferred to a memory module 160 in the system control 122. ~~When the scan is completed and an entire array of data has been acquired in the memory module 160, an~~ An array processor 161 operates to Fourier transform the data into an array of image data. This image data is conveyed through the serial link 115 to the computer system 107 where it is stored in the disk memory 111. In response to commands received from the operator console 100, this image data may be archived on the tape drive 112, or it may be further processed by the image processor 106 and conveyed to the operator console 100 and presented on the display 104.

[0030] For a more detailed description of the transceiver 150, reference is made to U.S. patent Nos. 4,952,877 and 4,922,736 which are incorporated herein by reference.

[0031] While many pulse sequences may be used to practice the present invention, in the preferred embodiment a 3D gradient-recalled echo pulse sequence is used to acquire the NMR data. Referring particularly to Fig. 2, an RF excitation pulse 220 having a flip angle of 45° is produced in the presence of a slab select gradient pulse 222 to produce transverse magnetization in the 3D volume of interest as taught in U.S. patent No. 4,431,968. This is followed by a phase encoding gradient pulse 224 directed along the z axis and a phase encoding gradient pulse 226 directed along the y axis. A readout gradient pulse 228 directed along the x axis follows and a partial echo (60%) NMR signal 230 is acquired and digitized as described above. After the acquisition, rewinder gradient pulses 232 and 234 are applied to rephase the magnetization before the pulse sequence is repeated as taught in U.S. patent No. 4,665,365.

[0032] As is well known in the art, the pulse sequence is repeated and the phase encoding pulses 224 and 226 are stepped through a series of values to sample the 3D k-space in the field of view. ~~As will be described in more detail below, it is the order in which these phase encoding gradients are stepped, and hence the order in which k-space is sampled, which is the subject of the present invention.~~ In the preferred embodiment 32 phase encodings are employed along the z axis and 256 phase encodings are employed along the y axis. Sampling along the

k_x axis is performed by sampling the echo signal 230 in the presence of the readout gradient pulse 228 during each pulse sequence. It will be understood by those skilled in the art that only a partial sampling along the k_x axis ~~is~~may be performed, and if this is done, the missing data is computed using a homodyne reconstruction or by zero filling. This enables the echo time (TE) of the pulse sequence to be shortened to 1.0 ms and the pulse repetition rate (TR) to be shortened to ~~4.5~~4.5 ms.

[0033] The preferred embodiment of the invention is a 3D CE-MRA scan of a subject after injection of a contrast agent. When the scan prescription is entered by the operator, scan parameters such as the imaging pulse sequence FOV and resolution are defined along all axes. Unique to the present invention, the FOV_{tot} is also defined. Based on this information a data matrix 16 is defined as illustrated in Fig. 4. Referring particularly to Fig. 5, after the injection of the contrast agent at process block 240, 2D images are rapidly acquired and reconstructed to display the vasculature at the starting boundary of the FOV_{tot} . Arrival of the contrast bolus is detected at process block 242 by observing when the arteries brighten. At this time table movement is initiated as indicated at process block 244. There are other methods for detecting bolus arrival, including automated methods such as that described in U.S. Pat. No. 6,167,293.

[0034] A loop is then entered at 246 in which MRI data is acquired as the table moves the patient through the sweet spot of the scanner. A table reference location is also established at this time. As indicated at process block 248, a view is acquired by performing the above-described pulse sequence. Many different view orders may be employed and the prescribed view order is stored in table 250. After acquisition of the NMR echo signal, the current table position is also acquired, as indicated at process block 252, and the NMR echo signal is Fourier transformed, along the x-axis as indicated at process block 254. As indicated at process block 256, the proper location in the data matrix 16 for the acquired view is then calculated and the data is stored therein as indicated at process block 258.

[0035] The location for each acquired view in the data matrix 16 is determined by two factors, the view number and the location of the patient table at the time the view was acquired. This is illustrated in Figs. 4 and 6 for a two-dimensional acquisition with a monotonic view order. The first view is acquired at echo location

$x(0)$ which serves as a reference location. Subsequent views are acquired as the y axis phase encoding is stepped through its 128 values. The location in the data matrix 16 along its k_y axis is determined by the $G_y G_x$ phase encoding for the view. As indicated by arrow 460, 260, the center of each successively acquired NMR echo signal after Fourier transformation along x is also shifted along the x-axis due to table movement. The amount of this shift from the reference position $x(0)$ is measured at the time of view acquisition using an electronic spatial encoder or by using the following equation:

$$\Delta = V_{ref} \cdot t$$

or alternatively

$$x(n) = V_{ref} \cdot n \cdot TR$$

where V_{ref} is the table velocity, t is the time past since the beginning of the scan, n is the number of pulse sequence repetitions since the start of the scan and TR is the time required for one pulse sequence repetition.

$$x(n) = V_{ref} \cdot t$$

or alternatively

$$x(n) = V_{ref} \cdot n \cdot TR$$

[0037]—where V_{ref} is the table velocity, t is the time past since the beginning of the scan, n is the number of pulse sequence repetitions since the start of the scan and TR is the time required for one pulse sequence repetition. The storage pattern is similar to a usual Cartesian k-space sampling pattern, but it is skewed by the table motion. When all the phase encoding views have been acquired, the process repeats and each new view of once transformed data is concatenated to the

previous acquired $k_x k_y$ view. It should be apparent that in a 3D acquisition a series of $k_z k_y$ phase encoded views are also acquired at each k_y phase encoding, and these are shifted and stored in a 3D data matrix 16 in the same manner.

~~[0036] [0038]—It can be appreciated that in general, the table motion might be such that the displacement of an acquired view along the x-axis X from one repetition to the next may is not be an integral number of X pixels. The data for such a view these repetitions must be shifted with a precision of less than one pixel.—~~
~~There are two possible methods with equivalent outcomes that can be applied to perform this data shift. The first called the direct method applies one phase shift to the raw, k-space data to shift it by to maintain consistency. The displacement $X(n)$ is known and can be split into two parts as follows:~~

~~$X(n) = m \cdot \Delta$. The second method called the time and memory efficient method uses a physical shift of the once Fourier transformed data plus an additional phase manipulation of the sampled raw k-space data.~~

~~[0037] [0039]— A. The direct method. $S_n(t)$ is recovered from $S'_n(t)$ by applying the phase term during demodulation of the raw continuous NMR signal.—~~
~~The process of demodulation of the raw signal as well as the application of phase accrual during the demodulation is well known to those studied in the art. The corrected NMR data is then sampled to give the same number of data points as are required for the number of pixels representing the FOV_{tot} in the X direction $x + \Delta x$ (example $N_{tot} = 2048$). After Fourier transform in X , the once-transformed hybrid-space signal is $H(x, k_y)$. The region representing FOV_x is retained by taking the product of $H(x, k_y)$ and a window function $W(x - \Delta)$ which has been shifted by a distance n)~~

(5)

~~where Δ . The simplest window function is the rectangle function $Rect(x)$, but other window functions which are non-zero outside FOV_x are possible. It is assumed that via the convolution theorem that convolution of the Fourier transform of $W(x)$ with $S_n(t)$ is equivalent to the product of $W(x)$ and $H(x, k_y)$ after Fourier transform of $S_n(t)$.—~~
~~Succeeding signals are added to the hybrid space matrix x is the pixel to pixel spacing along the x-direction:~~

$$\Delta x = \frac{FOV_x}{N_x} = \frac{FOV_{tot}}{N_{tot}} \quad (6)$$

where N_x and N_{tot} are the number of pixels in the X direction for the echo readout and for their specific k_y values after windowing to their specific known the total FOV respectively. Also, m is the largest integer such that $m \cdot \Delta$ and $FOV_x - x < X(n)$, and $\delta(n)$ is the sub-pixel-sized remainder which forces equation (5) to hold. As before, n is the index on the pulse sequence repetition number. The actual displacement positioning consists of two steps, coarse and fine, presented here in reverse order of application for ease of understanding.

[0040]— B. Time and memory efficient method. The signal $S'_n(t)$ is sampled for a specific FOV_x at some arbitrary FOV_y . The number of samples (N_x) is less than the number of pixels representing the final image (N_{tot}) with FOV_{tot} in the X direction. As an example $N_{tot} = 2048$, $N_x = 256$. Each pixel width represents a distance (D_{pix}) on the object where

Actual positioning of data in the Hybrid matrix requires two steps. They are presented in reverse order of application for ease of understanding.

[0038] [0041]— Step 2, coarse positioning: After the sampled k-space data is The Fourier transformed along x it can be signal is shifted by m pixels as given in the first term of equation (5) and directly placed into the hybrid matrix 16 with a positional accuracy of $\pm D_{pix} \Delta x$. The data then requires additional sub-pixel correction, by the amount $\delta(n)$ to maintain data consistency. This process is performed in step 1.

[0039] [0042]— Step 1, fine positioning: Prior to Fourier transform, transformation in X, a phase term twist representing the sub-pixel correction $\delta(n)$ is applied to along the sampled data. The subpixel shift (Δ') is given by:

$$\Delta' = \Delta - n \cdot D_{\text{pix}}$$

where n is chosen such that $\Delta' < D_{\text{pix}}$. The sampled signal with this fine correction is then given by:

where k is echo; i.e.

$$\hat{S}_n(k_x, k_y(n)) = e^{i2\pi \frac{k_x}{N_x} \frac{\delta(n)}{\Delta x}} \cdot S_n(k_x, k_y(n)) \quad (7)$$

Fourier transformation of \hat{S}_n along the number of the sampled data point ranging from 0 to $N_x - 1$. The sign applied to the phase term is dependent on whether the coarse correction will be a partial pixel less than or greater than the actual position.

Along the x-axis followed by the coarse positioning of Step 2 yields the desired $S'_n(x, k_y, (n))$.

In principle it would be possible to perform the entire displacement positioning using the phase manipulation approach of equation (7), but because multiple Fourier transformations of length N_{tot} would be required for each k_y value acquired, this "direct method" is not preferred because the two part position shifting and phase shifting has been found to be computationally more efficient.

[0040][0043]—It can be seen that after one complete cycle of k_y phase encoding values has been completed and stored in the data matrix 16, the process repeats itself and the phase encodings are applied a second time. Depending upon the number of phase encodings and the patient table velocity, the second time a view is acquired for a specific phase encoding, the placement of the echo signal samples in the data matrix 16 along the x direction may or may not overlap with some of the echo signal samples acquired for the previous measurement at that phase encoding. In the 2D example shown in Fig. 4, such a view is shown at $x(n)$ and the timing is such that there are no overlapping samples with the previously

acquired view at $x(0)$. In general some overlap will occur and the redundant, overlapping signal samples are averaged to improve SNR.

[0041][0044]—Referring again to Fig. 4, views are acquired and stored in the data matrix 16 until samples have been acquired for the entire extended field of view FOV_{tot} . The last view is indicated at 262 in Fig. 4, and after it is acquired and stored as determined at decision block 264, the scan is complete and a single image is reconstructed at process block 266 using the samples in data matrix 16. In the 3D acquisition of the preferred embodiment, this reconstruction includes Fourier transformation of the data matrix 16 along the k_y and k_z axes and calculation of pixel intensities from the resulting complex numbers. Typically, a two-dimensional projection image is then produced from the resulting three-dimensional array of image pixel intensities. The well-known maximum intensity pixel ("MIP") technique is preferred for producing the projection image and a pre-contrast mask image may also be subtracted to suppress signals from non-vascular tissues.

[0042][0045]—As shown in Fig. 4, the extended field of view FOV_{tot} over which complete k-space sampling is acquired is smaller than the x axis extent over which samples are acquired. The regions 268 and 270 of partial sampling at the beginning and end of the scan can also be reconstructed to extend the total field of view, but it can be appreciated that image quality will decline as fewer samples are used in the reconstruction.

[0043][0046]—Referring again to Fig. 5, after each view is acquired and stored during the scan, data in the data matrix 16 may be employed to reconstruct a two-dimensional monitor image as indicated at process block 272. Real-time images are thus produced for the operator who can use the information therein to control the scanning process. In addition to changing the usual scan parameters, the operator can adjust the table velocity as indicated at process block 274. This might be done, for example, to better match the speed at which the contrast bolus is moving through the extended field of view FOV_{tot} .

[0044][0047]—The real-time 2D monitor image may be produced in a number of ways. The preferred method is to reconstruct a full maximum intensity pixel (MIP) projection through the image stack as the data is being acquired. Unlike typical static 3D image acquisitions, a partial 3D data set can be reconstructed as

the hybrid matrix fills. For N_z slices and N_y phase encodes in Y, N_y Fourier transforms in the Z direction and N_z Fourier transforms in the Y direction can be performed with each TR. Each set of these described Fourier transform sets in Z and Y (including the previous transform in X) provide a partial 3D image set that is of one pixel length in the X direction. N_y pixels in the Y direction and N_z pixels in the Z direction. The maximum pixel value of a projection through this partial image set as well as earlier partial data sets can be found at the same time. The projection may be in the Z direction directly or in oblique angles through previously reconstructed partial image sets. The 2D MIP images can be displayed on the real-time monitor as each portion of the 3D data set is filled. Also, specific individual slices from the 3D data set can be viewed in real time rather than the MIP projection.

~~[0045][0048]—Alternatively the same process can be performed before the hybrid matrix is completely filled. No matter what phase encoding strategy is used, the phase encode order will periodically return to the center of k-space during the acquisition. Images viewed would possibly have a lower spatial resolution (in Y and Z) but would be more temporally recent than allowing full hybrid space filling. These real-time 2D images may be used to monitor scan progress. This not only includes the physical position of the acquisition on the patient at that time but can be used to monitor the contrast bolus itself. The bolus time-position information can be used to reduce or possibly increase the velocity of the scan during the runoff study.~~

[0046] These real-time 2D or low resolution 3D images may be used to monitor scan progress. This not only includes the physical position of the acquisition on the patient at that time but can be used to monitor the contrast bolus itself. The bolus time-position information can be used to reduce or possibly increase the velocity of the scan during the runoff study.

[0047][0049]—The table velocity to be used for the continuous motion is an independently chosen parameter. There are guidelines for its selection. In order to match table motion to the playout of phase encoding views the table should move a complete field-of-view (FOV_x , FOV_y) in the time required for application of one complete cycle of the phase encodings. At this velocity, consecutive measurements of the same view will abut and have neither an overlap ~~or~~ nor an intervening gap along the readout x direction. We define this as the “reference table velocity” and

for 2D imaging is given by:

$$V_{ref,2D} = \frac{FOV_x}{N_y \cdot TR} \left[\frac{cm}{sec} \right]$$

where N_y is the number of y phase encodes and TR is the repetition time of the imaging pulse sequence. For 3D imaging in order to allow for both phase encode directions the reference table velocity is given by:

$$V_{ref,3D} = \frac{FOV_x}{N_y \cdot N_z \cdot TR} \left[\frac{cm}{sec} \right]$$

where N_z is the number of z phase encodes. Note there is no dependence on the chosen resolution in X, only the chosen readout FOV_x . The table can be slowed or even stopped if redundant data are desired. In that case the redundant data can replace or be averaged with the non-zero data already stored in the data matrix 16.

~~[0048]~~ ~~[0050]~~ FOV_x can be variable during the scan. An acquisition can be performed with higher table velocity and thus greater temporal efficiency if a portion, or all of the hybrid space is acquired with a larger FOV_x . FOV_x can also be reduced if a higher spatial resolution is desired for a limited region within the FOV_{tot} . Regardless of changes in table speed or readout FOV_x , a single, data matrix 16 is formed from which a seamless image of the extended field of view FOV_{tot} can be reconstructed.

[0049] We have shown the feasibility of acquiring 2D and 3D images which cover longitudinal fields-of-view many times longer than the readout FOV possible in a single static image acquisition. The data for these images are acquired in conjunction with continuous table motion. This approach eliminates discontinuities in image signal at station boundaries as well as the need for overlap in FOV. The total scan times required for the moving table MRA acquisition vs. a generic fixed multi-station approach can be compared. For equal readout FOVs the continuous

motion approach of the present invention provides reduced overall scan time because time dedicated to station-to-station motion and dummy repetitions is eliminated.

[0050] With the introduction of shorter bore, less claustrophobic MRI systems, the present invention provides the ability to image large longitudinal fields of view in such systems. In addition, by using one of the new fast 3D steady state pulse sequences the invention allows whole body screening with one fast exam. This can be done using targeted contrast agents which determine, for example, the extent of metastases for certain cancers or the quantitative assessment of systemic thrombus.

[0051] In summary, continuously moving table acquisitions have been demonstrated for 2DFT and 3DFT MR acquisition using gradient echo and spin-echo methods. Phantom and in vivo images are sharp and appear devoid of motion artifacts as well as interference effects between consecutive FOVs. The method offers the potential for high resolution 2D and 3D MRI of extended objects.

[0052]

CLAIMS

1. In a magnetic resonance imaging (MRI) system having a defined field of view (FOV), a method for producing an image of a subject over an extended field of view (FOV_{tot}) which is larger than the FOV, the steps comprising:

- a) moving the subject through the MRI system such that the extended field of view (FOV_{tot}) passes through the defined field of view (FOV);
- b) continuously acquire NMR data from the subject as it is moved through the FOV by repeatedly performing an imaging pulse sequence which acquires NMR data comprising a view of the subject;
- c) adjusting each view acquired in step b) using subject position information;
- d) storing each adjusted view in a data matrix; and
- e) reconstructing an image using the data matrix.

2. The method as recited in claim 1 in which the MRI system has a table, and step a) is performed by:

- ~~i) placing the subject on the table; and~~
- ii)
 - i) placing the subject on the table; and
 - ii) moving the table.

3. The method as recited in claim 2 in which the table is moved continuously while performing step a).

5. The method as recited in claim 2 in which the table is moved at different velocities while performing step a).

6.

7. The method as recited in claim 2 which includes:

injecting the subject with a contrast agent; and

~~5. The method as recited in claim 2 which includes: injecting the subject with a contrast agent; and~~

in which the table is moved at a velocity which tracks the contrast agent as it moves through the extended field of view (FOV_{tot}).

8. The method as recited in claim 5 which includes:

~~6. The method as recited in claim 5 which includes: reconstructing monitoring images from the data matrix during the performance of step a) from data stored in the data matrix.~~

9. The method as recited in claim 1 in which step c) includes adjusting the location in the data matrix in which the view is stored in step d).

11. The method as recited in claim 1 in which step c) includes adjusting the phase of the NMR data in the view.

13. The method as recited in claim 2 in which step c) includes adjusting the location in the data matrix in which the view is stored in step d) as a function of the table location at the time the view is acquired in step b).

15. The method as recited in claim 2 in which step c) includes adjusting the phase of the NMR data in the view as a function of the table location at the time the view is acquired in step b).

17. The method as recited in claim 2 in which step b) further includes:

i) acquiring table location information as each view is acquired;

and

~~11. The method as recited in claim 2 in which step b) further includes:
i) acquiring table location information as each view is acquired;
and
the table location information is used in step c) to adjust each corresponding view.~~

18. The method as recited in claim 11 in which step c) includes:
i) performing a Fourier transformation of the NMR data in the view;
and

~~12. The method as recited in claim 11 in which step c) includes:
i) performing a Fourier transformation of the NMR data in the view; and
ii) calculating a location in the data matrix for the transformed view as a function of the table location at the time the view was acquired in step b).~~

19. The method as recited in claim 1 in which step c) includes:
i) adjusting the phase of the NMR data in the view;
ii) Fourier transforming the phase adjusted NMR data in the view; and
iii) adjusting the location in the data matrix in which the Fourier transformed view is stored in step d) as a function of subject location at the time the view is acquired in step b) with respect to a subject reference location.

21. The method as recited in claim 1 in which the performance of the imaging pulse sequence in step b) includes:

i) producing a readout magnetic field gradient during the acquisition of said NMR data comprising a view, and the readout magnetic field gradient is oriented in the same direction as subject movement.

23. The method as recited in claim 14 in which step c) includes:

i) Fourier transforming the acquired view; and
ii) adjusting the location in the data matrix in which the Fourier

transformed view is stored in step d) as a function of subject location at the time the view is acquired in step b) with respect to a subject reference location.

25. The method as recited in claim 1 in which the data matrix is a two-dimensional array of data.

27. The method as recited in claim 1 in which the data matrix is a three-dimensional array of data.

28.

29. In a magnetic resonance imaging (MRI) system, the improvement comprising:

- a) a table for supporting a subject and for moving the subject through a defined field of view (FOV) of the MRI system;
- b) a pulse generator for operating the MRI system under the direction of a pulse sequence to continuously acquire a series of NMR data views of the subject as the subject is moved through the FOV;
- c) means for adjusting each acquired view as a function of subject location at the time the view is acquired with respect to a reference subject location;
- d) a memory for storing the adjusted views as a data matrix; and
- e) means for reconstructing an image from data in the data matrix which has a field of view in the direction of table motion which is larger than the defined FOV.

30.

31. The improvement as recited in claim 18 in which element c) includes:

- i) means for Fourier transforming each acquired view; and
- ii) means for storing the Fourier transformed view in the data matrix at a location determined by the subject location at the time the view was acquired.

33. The improvement as recited in claim 18 which also includes:

- f) means for reconstructing an image from data in the data matrix as the subject is moved through the defined FOV and views are being acquired.

35. The improvement as recited in claim 20 which also includes:

- g) means for controlling the velocity of table motion as views are being acquired.

36.

ABSTRACT OF THE DISCLOSURE

MRA data is acquired from a large region of interest by translating the patient through the bore of the MRI system as a three-dimensional MRA data set are acquired. Patient table movement is controlled to track a bolus of contrast agent as it passes through the region of interest. Fluoroscopic images may be acquired during the scan to enable accurate bolus tracking. A seamless image of the entire region of interest is reconstructed.